

Sample Specification for Level Transmitter (Model 2852-LT)

The level probe shall be of a high frequency capacitance technology to monitor for the dielectric change as a liquid or solid rises or lowers in a tank or vessel.

A pulse card located at the probe will translate the probe signal to a frequency for high accuracy transmission via 2-conductor shielded cable. The control circuit can be mounted up to 1 km distance from the probe. The sensing probe shall be of Teflon and 316 SS wetted parts.

All calibration, power and control wiring shall be done at the wall mounted Type 4X/IP 65 control unit. Power input shall be specified as 24vdc or 80 to 240 vac.

The alarm relays shall be two SPDT, 10 amp dry contact. A time delay and sensitivity adjustment shall be standard to avoid nuisance alarms. A fully adjustable differential shall be available to allow the relay to cycle on and off between two distinct points on the probe.

A 4-20 mA non-isolated proportional level output signal will be available. An RS-485 Modbus communication port will be standard.

The probe length is to be sized to the application. For top mounting, determine the probe length required from the probe fitting to the probe tip to accommodate the lowest level measurement desired. .

The standard probe assembly shall carry a CSA Class 1 Div 1 Approval.

An optional Intrinsic Barrier may be specified to make the sensor assembly Intrinsically Safe where applicable (Part. #A00071)

The controller and probe shall be the Model 9852-LT as manufactured by Arjay Engineering Ltd., www.ArjayEng.com.